

- 6.1. The above concentration values and instrument conditions are for a Perkin-Elmer HGA- 2100, based on the use of a 20 μL injection, continuous flow purge gas and pyrolytic graphite. Smaller size furnace devices or those employing faster rates of atomization can be operated using lower atomization temperatures for shorter time periods than the above recommended settings.
- 6.2. The use of background correction is recommended.
- 6.3. Nitrogen may also be used as the purge gas.
- 6.4. For every sample matrix analyzed, verification is necessary to determine that method of standard addition is not required (see part 5.2.1 of the Atomic Absorption Methods section of this manual).
- 6.5. If method of standard addition is required, follow the procedure given earlier in part 8.5 of the Atomic Absorption Methods section of this manual.

7.0 Precision and Accuracy

- 7.1. Precision and accuracy data are not available at this time.