

5.0 Interference

- 5.1 Interferences from sulfur compounds are eliminated by acidifying the sample to a pH of less than 4 with H_3PO_4 and aerating briefly by stirring and adding CuSO_4 .
- 5.2 Oxidizing agents such as chlorine, detected by the liberation of iodine upon acidification in the presence of potassium iodide, are removed immediately after sampling by the addition of an excess of ferrous ammonium sulfate (7.10). If chlorine is not removed, the phenolic compounds may be partially oxidized and the results may be low.

6.0 Apparatus

- 6.1 Distillation apparatus, all glass consisting of a 1 liter pyrex distilling apparatus with Graham condenser.
- 6.2 pH meter.
- 6.3 Spectrophotometer, for use at 460 or 510 nm.
- 6.4 Funnels.
- 6.5 Filter paper.
- 6.6 Membrane filters.
- 6.7 Separatory funnels, 500 or 1,000 mL.
- 6.8 Nessler tubes, short or long form.

7.0 Reagents

- 7.1 Phosphoric acid solution, 1 + 9: Dilute 10 mL of 85% H_3PO_4 to 100 mL with distilled water.
- 7.2 Copper sulfate solution: Dissolve 100 g $\text{CuSO}_4 \cdot 5\text{H}_2\text{O}$ in distilled water and dilute to 1 liter.
- 7.3 Buffer solution: Dissolve 16.9 g NH_4Cl in 143 mL conc. NH_4OH and dilute to 250 mL with distilled water. Two mL should adjust 100 mL of distillate to pH 10.
- 7.4 Aminoantipyrine solution: Dissolve 2 g of 4AAP in distilled water and dilute to 100 mL.
- 7.5 Potassium ferricyanide solution: Dissolve 8 g of $\text{K}_3\text{Fe}(\text{CN})_6$ in distilled water and dilute to 100 mL.
- 7.6 Stock phenol solution: Dissolve 1.0 g phenol in freshly boiled and cooled distilled water and dilute to 1 liter. 1 mL = 1 mg phenol.
- 7.7 Working solution A: Dilute 10 mL stock phenol solution to 1 liter with distilled water. mL = 10 μg phenol.
- 7.8 Working solution B: Dilute 100 mL of working solution A to 1000 mL with distilled water. 1 mL = 1 μg phenol.
- 7.9 Chloroform
- 7.10 Ferrous ammonium sulfate: Dissolve 1.1 g ferrous ammonium sulfate in 500 mL distilled water containing 1 mL conc. H_2SO_4 and dilute to 1 liter with freshly boiled and cooled distilled water.

8.0 Procedure

- 8.1 Distillation

- 8.1.1 Measure 500 mL sample into a beaker. Lower the pH to approximately 4 with 1 + 9 H₃PO₄ (7.1), add 5 mL CuSQ solution (7.2) and transfer to the distillation apparatus. Omit adding H₂PO₄ and CuSO₄ if sample was preserved as described in 4.1.
- 8.1.2 Distill 450 mL of sample, stop the distillation, and when boiling ceases add 50 mL of warm distilled water to the flask and resume distillation until 500 mL have been collected.
- 8.1.3 If the distillate is turbid, filter through a prewashed membrane filter.
- 8.2 Direct photometric method
- 8.2.1 Using working solution A (7.7), prepare the following standards in 100 mL volumetric flasks.

mL of working solution A	Conc. $\mu\text{g/L}$
0	0.0
0.5	50.0
1.0	100.0
2.0	200.0
5.0	500.0
8.0	800.0
10.0	1000.0

- 8.2.2 To 100 mL of distillate or an aliquot diluted to 100 mL and/or standards, add 2 mL of buffer solution (7.3) and mix. The pH of the sample and standards should be 10 ± 0.2 .
- 8.2.3 Add 2.0 mL aminoantipyrine solution (7.4) and mix.
- 8.2.4 Add 2.0 mL potassium ferricyanide solution (7.5) and mix.
- 8.2.5 After 15 minutes read absorbance at 510 nm.
- 8.3 Chloroform extraction method
- 8.3.1 Using working solution B (7.8), prepare the following standards. Standards may be prepared by pipetting the required volumes into the separatory funnels and diluting to 500 mL with distilled water.

mL of working solution B	Conc. $\mu\text{g/L}$
0.0	0.0
3.0	6.0
5.0	10.0
10.0	20.0
20.0	40.0
25.0	50.0

- 8.3.2 Place 500 mL of distillate or an aliquot diluted to 500 mL in a separatory funnel. The sample should not contain more than 25 μg phenol.
- 8.3.3 To sample and standards add 10 mL of buffer solution (7.3) and mix. The pH should be 10 ± 0.2 .
- 8.3.4 Add 3.0 mL aminoantipyrine solution (7.4) and mix.
- 8.3.5 Add 3.0 mL potassium ferricyanide solution (7.5) and mix.
- 8.3.6 After three minutes, extract with 25 mL of chloroform (7.9). Shake the separatory funnel at least 10 times, let CHCl₃ settle, shake again 10

- times and let chloroform settle again. Vent chloroform fumes into hood.
- 8.3.7 Filter chloroform extracts through filter paper. Do not add more chloroform. Carry out filtration in a hood. Dispose of chloroform in environmentally acceptable manner.
 - 8.3.8 Read the absorbance of the samples and standards against the blank at 460 nm.

9.0 Calculation

- 9.1 Prepare a standard curve by plotting the absorbance value of standards versus the corresponding phenol concentrations.
- 9.2 Obtain concentration value of sample directly from standard curve.

10.0 Precision and Accuracy

- 10.1 Using the extraction procedure for concentration of color, six laboratories analyzed samples at concentrations of 9.6, 48.3, and 93.5 $\mu\text{g/L}$. Standard deviations were ± 0.99 , ± 3.1 and ± 4.2 $\mu\text{g/L}$, respectively.
- 10.2 Using the direct photometric procedure, six laboratories analyzed samples at concentrations of 4.7, 48.2 and 97.0 mg/L . Standard deviations were ± 0.18 , ± 0.48 and ± 1.58 mg/L , respectively.

Bibliography

1. Annual Book of ASTM Standards, Part 31, "Water", Standard D 1783-70, p553 (1976).
2. Standard Methods for the Examination of Water and Wastewater, 14th Edition, p574-581, Method 510 through 510C, (1975).